

# Service Repair Sheet SRS68994

Contact Name  
Company/ Hospital Name  
Department  
Position  
Direct Phone  
General Phone  
Opera Account  
Email  
Order Number  
Date Received  
Booked in By  
Main Company  
Type Return

Charlie Brown  
St James University Hospital  
Clinical Engineering  
Senior/Lead Technician  
0113 2064288  
01132433144  
00002570  
Charlie.brown@nhs.net  
  
14/Jan/2025  
Robert Connor  
Viamed  
Quote

## VIAMEDclean

**Goods In Only**  
Decontamination  
certificate provided  
by customer ☒

Cleaned by Viamed,  
if no declaration certificate  
from customer ☒  
Signed: \_\_\_\_\_  
Date: \_\_\_\_\_

## Goods Out Only

Cleaned by Viamed before  
returning to customer  
Signed: \_\_\_\_\_  
Date: \_\_\_\_\_

Notes 07/Jan/2025 Steve Hardaker

07/Jan/2025 Steve Hardaker

Is returning a Tom Thumb TT490-15 for a service and 3m hose replacement.

Serial number: VMCA08

08/Jan/2025 Steve Hardaker

Sending this week, needs quick turnaround, we have 1x 0330218 hose in stock.

14/Jan/2025 Robert Connor

Received 1 x tom thumb s/n VMCA08

Ready For quote

C. Green 15-1-25

Repair Complete Signed

C. Green 30-1-25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37577	Tom Thumb	0310030	VMCA08	N

0380000 x 1 @ £90  
S/N, SRS, SRN

0330218 x 1 @ £44.50  
~~S/N~~ SRS, SRN

UPS x 1 £12