

Viamed Limited · 15 Station Road · Cross Hills Keighley · West Yorkshire · BD20 7DT · United Kingdom Tel: +44 (0)1535 634 542 · Fax: +44 (0)1535 635 582 Email: info@viamed.co.uk Website: www.viamed.co.uk

Service Repair Sheet 68994

Contact Name

Charlie Brown

Company/ Hospital Name

St James University Hospital

Department

Clinical Engineering

Position Direct Phone Senior/Lead Technician 0113 2064288

General Phone

01132433144

Opera Account

00002570

Email

Charlie.brown@nhs.net

Order Number

RR8N400547119

Date Received Booked in By

14/Jan/2025 Robert Connor

Repair

S/N Ref SRN37577 0310034 VMCA08 **Equipment Type** 

Tom Thumb

Complete - Repaired Time: 0 Hour(s)

Parts Replaced

0380000 Qty: 1

Qty: 1 0330218

Tom Thumb TT490 VMCA08 has been stripped and cleaned then reassembled with new O rings and a new oxygen hose. The two valves have been checked and reset before QA.