

Service Repair Sheet 69000

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Diane Cobbledick

Wansbeck General Hospital

Medical Electronics

Senior Medical Electronics Technician

01670529020

01670521212

00000161

Diane.Cobbledick@northumbria-healthcare.nhs.uk

RTFE400410108

16/Jan/2025

Robert Connor

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN37582	1410000	PR03124A12	Foetal Simulator	Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.