



Service Repair Sheet 69024

Contact Name

Company/ Hospital Name

Department

Position

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Richard Tuckwood

Nottingham City Hospital

Medical Equipment Services Unit (MESU)

Team Leader

01159691169

00003910

Richard.Tuckwood@nuh.nhs.uk

202013674

06/Feb/2025

Robert Connor

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN37601	5610152	AL000102	Blender	Time :0 Hour(s)

Parts Replaced

Qty: 1 0380050

Qty: 1 0332004

Blender AL000102 has been stripped and ultrasonically cleaned. It has been rebuilt, using the service kit, and tested as per the service manual.