

Viamed Limited - 15 Station Road Cross Lines Keighley · West Yorkshire · BD20 7DT · United Kingdom Tel: +44 (0)1535 634 542 · Fax: +44 (0)1535 635 582 Email: info@viamed.co.uk Website: www.viamed.co.uk

Service Repair Sheet 69005

Contact Name

Company/ Hospital Name

Position **Direct Phone** General Phone Opera Account

Email

**Date Received** Booked in By

**Debbie Summers** 

Elham Valley Referrals Head Veterinary Nurse

01303 840499 01303 8490499 00012494

debbie@elhamvalley.com

20/Jan/2025 Robert Connor

Repair SRN37583

S/N 2510011 M0005639 **Equipment Type** 

Complete - Repaired Microstim DB3 Time: 0 Hour(s)

Parts Replaced

2540010 Qty: 1 2520000 Qty: 1

The Microstim DB3 has been serviced and the patient lead has been replaced.