

Service Repair Sheet 69029

| | |
|------------------------|--------------------------------------|
| Contact Name | Owen Daniel |
| Company/ Hospital Name | Royal Albert Edward Infirmary |
| Department | Medical Electronics / ELS Department |
| Position | Medical Engineer |
| Direct Phone | 01942 822002 |
| General Phone | 01942 24 40 00 |
| Opera Account | 00005363 |
| Email | Owen.Daniels@wwl.nhs.uk |
| Date Received | 11/Feb/2025 |
| Booked in By | Kate Griffiths |

| | | | | |
|----------|---------|----------|----------------|---------------------|
| Repair | Ref | S/N | Equipment Type | Complete - Repaired |
| SRN37625 | 0310200 | B1301409 | Blender | Time :0 Hour(s) |

Parts Replaced

| | |
|--------|---------|
| Qty: 1 | 0380019 |
| Qty: 1 | 0332001 |

Blender B1301409 has been stripped and cleaned. All O rings, diaphragms etc have been replaced and it has been calibrated.
