

Service Repair Sheet 69022

Contact Name	Paul Doran
Company/ Hospital Name	Hexham General Hospital
Direct Phone	01434655403
General Phone	01434 655655
Opera Account	00002220
Email	paul.doran@northumbria-healthcare.nhs.uk
Order Number	RTFE400414442
Date Received	10/Feb/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37605	1410000	PR03124A11	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1430309		
	Qty: 1	1480000		

V1000 has had transducer interface cushion replaced and calibration has been checked.
