

Equipment for Repair

Repair Reference - 77650

Date: 19/2/25
To: VIANCO

Vet Direct
Units 9 & 10, Gateway West
Newburn Riverside
Newcastle upon Tyne
NE15 8NZ

Tel: 0191 2642705
Email: repairs@vet-direct.com

The enclosed product has been sent to you for inspection. Please provide estimate to repair - please send estimate to repairs@vet-direct.com

Product:	ENVITE OXIGANT S		
Serial No:	A111116 / A106327		
Accessories:	Full Box		
Quantity:	1		
Condition:	USED		
Work required:	UNABLE TO GET BELOW 20.9		
Warranty:	<input checked="" type="radio"/> Yes / <input type="radio"/> No	14/8/24	PO Number: V01533
Quotation required?	Yes / No		
Supplier action required:	<input checked="" type="radio"/> Repair / <input type="radio"/> Replace / <input type="radio"/> Credit		