

Service Repair Sheet SRS69038

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Derek Smith
Vet Direct Services Ltd
Repairs Department
Technical Support Co-Ordinator

01912642705
00012824
derek.smith@vet-direct.com

25/Feb/2025
Robert Connor
Viamed
Warranty

VIAMEDclean
Goods In Only
 Decontamination certificate provided by customer ☒

Cleaned by Viamed, if no declaration certificate from customer ☒
 Signed: RS
 Date: 23/2/25

Goods Out Only
 Cleaned by Viamed before returning to customer
 Signed: _____
 Date: _____

Notes 17/Feb/2025 Robert Connor
25/Feb/2025 Robert Connor

Received 1 x OxiQuant S s/n A106327, with OOI103 s/n A111116, with flow diverter, blue tee, and 2 x oxygen stem.

Ready For quote

Repair Complete Signed

CGreen 28.2.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37628	Oxygen Analyser	0111275	A106327	Yes
SRN37629	Oxygen Sensor	0110660	A111116	<u>r</u>

Part of unit
cg 28.2.25.

0140137 x 1 FOC
SRS, SRN
S/N A106327 SRS 69038 SRN 37628
Repaired under warranty

0140011 x 1 FOC
S/N, SRS, SRN
No fault found.

UPS > 1 FOC.