

Service Repair Sheet SRS69011

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Tim Holland

North Manchester General Hospital

Clinical Engineering Dept.

Senior Clinical Engineer

01617202347

01617954567

00003580

timothy.holland@mft.nhs.uk

29/Jan/2025

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 24/Jan/2025 Steve Hardaker

24/Jan/2025 Steve Hardaker

Returning a Yom Thumb he received back from SRS 68974, it doesn't register a pressure until it gets to over 10lpm, sn 040463.

Also returning another one for for service sn 040466.

29/Jan/2025 Robert Connor

Received 2 x tom thumb s/n 040463, 040466

Ready For quote

CGreen 30-1-25

Repair Complete Signed

~~CGreen 30-1-25~~
CGreen 5-3-25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37596	Tom Thumb	0310030	040463	Yes (r)
SRN37597	Tom Thumb	0310030	040466	

last serviced Jan 18/1/25

0340000 x 1 FOC → SRN 37596.

S/N, SRN, SRS

Evaluated under repair warranty

0380000 x 1 @ £90 → SRN 37597

S/N, SRS, SRN

UPS x 1 @ £12