

Service Repair Sheet 69011

Contact Name	Tim Holland
Company/ Hospital Name	North Manchester General Hospital
Department	Clinical Engineering Dept.
Position	Senior Clinical Engineer
Direct Phone	01617202347
General Phone	01617954567
Opera Account	00003580
Email	timothy.holland@mft.nhs.uk
Order Number	000494432
Date Received	29/Jan/2025
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN37596	0310030	040463	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Steve Hardaker and myself have stripped and thoroughly inspect this TomThumb.

It has been tested down to 5lpm and passed QA.

---

Repair	Ref	S/N	Equipment Type	
SRN37597	0310030	040466	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced	
	Qty: 1    0380000

Tom Thumb TT480 040466 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

---