



Kate Griffiths &lt;viamed.kate.griffiths@gmail.com&gt;

---

**Fwd: Oxygen Monitor / Analyser**

1 message

---

**Steve Hardaker** <steve.hardaker@viamed.co.uk>  
To: office@vmsecure.me.uk

7 March 2025 at 12:49

Please can someone send a quotation to Lewis for:

**5x 0111277 - Oxygen monitor MySign O Kit @ £405.20+VAT**

Thanks,  
Steve H

----- Forwarded message -----

From: **BRAZIER, Lewis (EAST KENT HOSPITALS UNIVERSITY NHS FOUNDATION TRUST)** <lewis.brazier@nhs.net>  
Date: Fri, 7 Mar 2025 at 11:47  
Subject: RE: Oxygen Monitor / Analyser  
To: Steve Hardaker <steve.hardaker@viamed.co.uk>

Hi Steve,

If we could get the one with the official charger please.

**If any collections are being arranged please book these for a weeks time, as I do not always work onsite and may miss collections dates.**

**Please Note my non-technical working / limited-contactability days are:**

-----

**Please Note that my next Planned Leave is:**

**7<sup>th</sup> March – 22<sup>nd</sup> March**

-----

Your feedback is invaluable to us - If you have a moment, please share your thoughts to help us improve:  
<https://forms.office.com/e/UFm3sGyxm6>

**Thank you**

**Lewis Brazier**

Senior Electrical Medical Engineering Technician, 2gether Support Solutions

DD: 01233 616601 | Ext: 723-6601 | Mob: N/A | [Lewis.Brazier@nhs.net](mailto:Lewis.Brazier@nhs.net)Equipment Management & Engineering (EME), William Harvey Hospital,  
Kennington Road, Ashford, Kent, TN24 0LZ

My Working Days are Monday – Friday, Hours 0830 - 1630



2gether Support Solutions Limited is a company registered in England and Wales with Company Number 11385580.  
Registered Office: Management Offices, William Harvey Hospital, Kennington Road, Willesborough, Ashford, Kent, England,  
TN24 0LZ. VAT No. GB 297 8726 29.



Please consider your environmental responsibility. Before printing this e-mail message, ask yourself whether you really need a hard copy.

**From:** Steve Hardaker <[steve.hardaker@viamed.co.uk](mailto:steve.hardaker@viamed.co.uk)>

**Sent:** 07 March 2025 11:17

**To:** BRAZIER, Lewis (EAST KENT HOSPITALS UNIVERSITY NHS FOUNDATION TRUST) <[lewis.brazier@nhs.net](mailto:lewis.brazier@nhs.net)>

**Subject:** Re: Oxygen Monitor / Analyser

You don't often get email from [steve.hardaker@viamed.co.uk](mailto:steve.hardaker@viamed.co.uk). [Learn why this is important](#)

This message originated from outside of NHSmail. Please do not click links or open attachments unless you recognise the sender and know the content is safe.

Good morning Lewis,

Do you want the kit with the official charger, or the basic unit? Both come with a sensor and T-piece, the only difference is the inclusion of a charger.

The MySign can be charged in a USB-C socket on a PC, but if that's a 500mA port, it may take up to 6 hours. The charger is a USB-C 1A charger, so the charging time will be around half the time of using a PC.

Alternatively, you can use a mobile phone charger, **but not if the device is going to be used in a patient environment**, as the charger must be accredited for that application, which the official one is.

**p/n 0111276 - Oxygen monitor MySign O @ £356.00+VAT**

**p/n 0111277 - Oxygen monitor MySign O Kit @ £405.20+VAT**

MySignO Oxygen Monitor – Viamed Ltd

If you can let me know, I will get a formal quotation raised, thanks.

Best regards,

Steve Hardaker  
Technical Support Manager  
Viamed Ltd.

**Please note: Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office. Telephone calls to the main office will be answered, but please continue to use email where possible.**

Website: [www.viamed.co.uk](http://www.viamed.co.uk)  
email: [steve.hardaker@viamed.co.uk](mailto:steve.hardaker@viamed.co.uk)  
Tel: +44 (0)1535 634542  
Fax: +44 (0)1535 635582

Sign up to our mailing list, go to [www.viamed.co.uk/sign-up](http://www.viamed.co.uk/sign-up)  
Twitter: [twitter.com/ViamedLtd](https://twitter.com/ViamedLtd)  
Facebook: [facebook.com/ViamedLtdUK](https://facebook.com/ViamedLtdUK)  
Company registered in England, No. 1291765.

E-mail Disclaimer

-----  
The information in this email is confidential and may be legally privileged.  
It is intended solely for the addressee. Access to this email by anyone else is unauthorized.  
If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542  
GDPR / Privacy Policy available at [www.viamed.co.uk/policy](http://www.viamed.co.uk/policy)  
-----

On Fri, 7 Mar 2025 at 11:05, BRAZIER, Lewis (EAST KENT HOSPITALS UNIVERSITY NHS FOUNDATION TRUST) <[lewis.brazier@nhs.net](mailto:lewis.brazier@nhs.net)> wrote:

Good morning,

Would I be able to get a quote for 5 MySign O's with the quickest reaction sensor and T piece please.

**If any collections are being arranged please book these for a weeks time, as I do not always work onsite and may miss collections dates.**

**Please Note my non-technical working / limited-contactability days are:**

-----  
**Please Note that my next Planned Leave is:**

**7<sup>th</sup> March – 22<sup>nd</sup> March**  
-----

Your feedback is invaluable to us - If you have a moment, please share your thoughts to help us improve:  
<https://forms.office.com/e/UFm3sGyxm6>

**Thank you**

**Lewis Brazier**

Senior Electrical Medical Engineering Technician, 2gether Support Solutions

DD: 01233 616601 | Ext: 723-6601 | Mob: N/A | [Lewis.Brazier@nhs.net](mailto:Lewis.Brazier@nhs.net)Equipment Management & Engineering (EME), William Harvey Hospital,  
Kennington Road, Ashford, Kent, TN24 0LZ

My Working Days are Monday – Friday, Hours 0830 - 1630



2gether Support Solutions Limited is a company registered in England and Wales with Company Number 11385580.  
Registered Office: Management Offices, William Harvey Hospital, Kennington Road, Willesborough, Ashford, Kent, England,  
TN24 0LZ. VAT No. GB 297 8726 29.



Please consider your environmental responsibility. Before printing this e-mail message, ask yourself whether you really need a hard copy.

**From:** Steve Hardaker <[steve.hardaker@viamed.co.uk](mailto:steve.hardaker@viamed.co.uk)>**Sent:** 04 March 2025 15:36**To:** BRAZIER, Lewis (EAST KENT HOSPITALS UNIVERSITY NHS FOUNDATION TRUST) <[lewis.brazier@nhs.net](mailto:lewis.brazier@nhs.net)>**Subject:** Re: Oxygen Monitor / AnalyserYou don't often get email from [steve.hardaker@viamed.co.uk](mailto:steve.hardaker@viamed.co.uk). [Learn why this is important](#)

This message originated from outside of NHSmail. Please do not click links or open attachments unless you recognise the sender and know the content is safe.

Good afternoon Lewis,

Thank you for your enquiry concerning oxygen monitors.

For the application you described, either the MySign O or the MaxO2 ME would be suitable, but I would advise against the UltraMax O2 unless you need a device specifically for testing oxygen concentrators, as it uses an ultrasonic sensor that is calibrated for the output from a concentrator and not medical grade air or oxygen (it won't read above 96%).

Both devices are similar spec, with a T90\* response time of <12s for the MySign O and <15s for the MaxO2 ME.

\*Note: this is the time taken to cover 90% of a change between different oxygen levels and can be used as a benchmark, it's not the time taken to get full accuracy as the rate of change slows down as it approaches the final reading, like a flattening off curve if you imagine it on a graph.

The MySign O is a rechargeable device with an operating life of around 2 days continuous use per charge. It can be operated continuously from a mains power adapter if you purchase the kit option detailed below:

**p/n 0111276 - Oxygen monitor MySign O @ £356.00+VAT**

**p/n 0111277 - Oxygen monitor MySign O Kit @ £405.20+VAT**

[MySignO Oxygen Monitor – Viamed Ltd](#)

If you wanted something capable of running continuously for months on batteries, then the MaxO2 ME would be a better option, with up to 5,000 hours use on a set of AA batteries. The display isn't quite as nice, it's just a backlit LCD versus a colour TFT, so it really depends on what matters more, a long battery life or a clearer display.

**p/n 0111263 - MaxO2 ME Oxygen monitor @ £619.70+VAT**

[MaxO2 ME Oxygen Monitor – Viamed Ltd](#)

The MySign O has an in-built hypoxia alarm that triggers below 18%, which is the minimum level for the low alarm limit. The MaxO2 ME can be set as low as 15% if required.

Both devices feature an integral kick-stand, but other mounting options are available, such as wall-mounts, pole mounts and rail mounts. These are detailed in the attached oxygen monitoring catalogue.

I have attached the product leaflets for reference. Please can you review and if you have any questions, or if you need a quotation, please let me know.

Best regards,

Steve Hardaker  
Technical Support Manager  
Viamed Ltd.

**Please note: Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office. Telephone calls to the main office will be answered, but please continue to use email where possible.**

Website: [www.viamed.co.uk](http://www.viamed.co.uk)  
email: [steve.hardaker@viamed.co.uk](mailto:steve.hardaker@viamed.co.uk)  
Tel: +44 (0)1535 634542  
Fax: +44 (0)1535 635582

Sign up to our mailing list, go to [www.viamed.co.uk/sign-up](http://www.viamed.co.uk/sign-up)  
Twitter: [twitter.com/ViamedLtd](https://twitter.com/ViamedLtd)  
Facebook: [facebook.com/ViamedLtdUK](https://facebook.com/ViamedLtdUK)  
Company registered in England, No. 1291765.

E-mail Disclaimer

-----  
The information in this email is confidential and may be legally privileged.  
It is intended solely for the addressee. Access to this email by anyone else is unauthorized.  
If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542  
GDPR / Privacy Policy available at [www.viamed.co.uk/policy](http://www.viamed.co.uk/policy)  
-----

Good morning,

I am a Senior Technician for Medical engineering at William Harvey Hospital.

I have been asked to look into getting a new engineering standard Oxygen Analyser. I have looked at your website and I really like:



I especially like the look of the MySignO. I would be looking at purchasing 3 units if possible. We need to be able to monitor Oxygen level outputs of a varied medical device model base. We are also wanting quick reaction oxygen sensors, as we run a lot of our tests off of bottles I would like to limit the amount of wastage.

Is this something you would be able to support us with?

**If any collections are being arranged please book these for a weeks time, as I do not always work onsite and may miss collections dates.**

**Please Note my non-technical working / limited-contactability days are:**

-----

**Please Note that my next Planned Leave is:**

**7<sup>th</sup> March – 22<sup>nd</sup> March**

-----

Your feedback is invaluable to us - If you have a moment, please share your thoughts to help us improve: <https://forms.office.com/e/UFm3sGyxm6>

**Thank you**

**Lewis Brazier**

Senior Electrical Medical Engineering Technician, 2gether Support Solutions

DD: 01233 616601 | Ext: 723-6601 | Mob: N/A | [Lewis.Brazier@nhs.net](mailto:Lewis.Brazier@nhs.net)Equipment Management & Engineering (EME), William Harvey Hospital,  
Kennington Road, Ashford, Kent, TN24 0LZ

My Working Days are Monday – Friday, Hours 0830 - 1630



\*\*\*\*\*  
\*\*\*\*\*

This message may contain confidential information. If you are not the intended recipient please:

i) inform the sender that you have received the message in error before deleting it; and  
ii) do not disclose, copy or distribute information in this e-mail or take any action in relation to its content (to do so is strictly prohibited and may be unlawful).  
Thank you for your co-operation.

NHSmial is the secure email, collaboration and directory service available for all NHS staff in England. NHSmial is approved for exchanging patient data and other sensitive information with NHSmial and other accredited email services.

For more information and to find out how you can switch visit [Joining NHSmial – NHSmial Support](#)

\*\*\*\*\*  
\*\*\*\*\*

This message may contain confidential information. If you are not the intended recipient please:

i) inform the sender that you have received the message in error before deleting it; and  
ii) do not disclose, copy or distribute information in this e-mail or take any action in relation to its content (to do so is strictly prohibited and may be unlawful).  
Thank you for your co-operation.

NHSmial is the secure email, collaboration and directory service available for all NHS staff in England. NHSmial is approved for exchanging patient data and other sensitive information with NHSmial and other accredited email services.

For more information and to find out how you can switch visit [Joining NHSmial – NHSmial Support](#)

---

**2 attachments**



image002.jpg  
40K



image004.jpg  
8K