

Service Repair Sheet 69048
Contact Name Dave Bell
Company/ Hospital Name North Tyneside General Hospital
Direct Phone 03448118111
Opera Account 00003859
Email david.bell2@NHCT.NHS.UK
Order Number RTFE400417239
Date Received 26/Feb/2025
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37652	1410000	PR03124A10	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
