

Service Repair Sheet SRS69047

Contact Name Sudath Weeraman
 Company/ Hospital Name Prime Lanka Associate
 Department
 Position
 Direct Phone 9410718732285
 General Phone
 Opera Account 00006714
 Email sudathweeraman17@gmail.com
 Order Number
 Date Received 13/Mar/2025
 Booked in By Robert Connor
 Main Company Viamed
 Type Return Warranty

VIAMEDclean	
Goods In Only	
Decontamination certificate provided by customer	<input checked="" type="checkbox"/>
Cleaned by Viamed, If no declaration certificate from customer	<input type="checkbox"/>
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 25/Feb/2025 Kate Griffiths

25/Feb/2025 Kate Griffiths

1 x 0110429 to send in as faulty under warranty. S/n: KH39399174. Fault - varying oxygen percentage. Set 100% indicate 30%

13/Mar/2025 Robert Connor

Received 1 X MAX-250E s/n KH39399174

Ready For quote

Repair Complete Signed

C Green 14.3.25

GBP
Dist.

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37686	Oxygen Sensor	0110429	KH39399174	Yes

0146013 x 1 FOC.

Old S/N, SRS, SRN

0110429 x 1 FOC

S/N, SRS, SRN

EXW. 23x12x12cm 0.30kg. to include SRS 69053 as well.