

Requested RMA
21-2-25

Service Repair Sheet SRS69026

Contact Name	Mick Smithson
Company/ Hospital Name	Lamboo Medical (UK)
Department	
Position	Duty Manager
Direct Phone	0194 2724272
General Phone	01942 724 272
Opera Account	00013111
Email	Mick@lamboo-medical.com
Order Number	
Date Received	17/Feb/2025
Booked in By	Robert Connor
Main Company	Viamed
Type Return	Warranty

VIAMEDclean	
Goods In Only Decontamination certificate provided by customer	<input type="checkbox"/>
Cleaned by Viamed, if no declaration certificate from customer	<input checked="" type="checkbox"/>
Signed: <u>RC</u>	
Date: <u>17/2/25</u>	
Goods Out Only Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 07/Feb/2025 Steve Hardaker

07/Feb/2025 Steve Hardaker

Returning 2x MySign O under warranty, s/n 103629 and 103633, both displaying error code

07/Feb/2025 Steve Hardaker

Returning 2x MySign O under warranty, s/n 103629 and 103633, both displaying error code "Hardware 000 180"
These will need to go to Envitec.

Please bring to the attention of Steve Nixon as he may want to discuss with Envitec whether this is a known problem.

14/Feb/2025 Steve Hardaker

Shipped today, so should be arriving Monday 17/2/25.

17/Feb/2025 Robert Connor

Received 2 x MySign O s/n 103629, 103633

Ready For quote

Repair Complete Signed

CAneen 17.3.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37622	Oxygen Monitor	0111276	103629	Yes
SRN37623	Oxygen Monitor	0111276	103633	Yes



0140138 x 2 FOC
S/N, SRS, SRN

UPS x 1 @ £0.