

Box 1023
emailed Supplier
4-3-25

Service Repair Sheet SRS69021

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Lisa Angell

Royal Veterinary College

01707 669221

01707 666 366 (Reception)

00012213

langell@rvc.ac.uk

25/Feb/2025

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 04/Feb/2025 Steve Hardaker

04/Feb/2025 Steve Hardaker

Sending a veterinary VM-2500 that has stopped working. S/n 81040139, sold 2015. Says it won't power up and wants a repair quote.

19/Feb/2025 Steve Hardaker

She had some success turning it on but it went straight back off again with the rechargeable battery in it, and strangely she said it doesn't work with AA batteries. Is sending it in.

25/Feb/2025 Robert Connor

Received 1 x VM-2500-S-Vet s/n 81040139, with rev 1 power supply and UK adapter.

Ready For quote

C.Green 18-3-25

Repair Complete Signed _____

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37626	Capnograph	4410540	81040139	N

4440000 x 1 @ £95 per hr = 1/2 hr charge £47.50
S/N, SRS, SRN.

4480520 x 1 @ £160
SRS, SRN

4430120 x 1 @ £148
SRS, SRN

4400070 x 1 @ £51.60
SRS, SRN

4430201 x 1 @ £15.40
SRS, SRN

4430100 x 1 @ £90
SRS, SRN. service exchange.

UPS x 1 @ £12.