

Declaration of contamination status

From (consignor):	Lisa Angell	To (consignee):	Viamed
Address:	QMHA, Hawkshead Lane	Address:	Returns department
	AL97TA		15 Station Road, BD20 7DT
Reference:		Reference:	SRS69021
Emergency tel:			

Type of equipment:	Monitoring device	Manufacturer:	Viamed
Description of equipment:	Portable Capnograph and pulse oximeter		
Other identifying marks:			
Model No:	VM-2500-S	Serial No:	81040139
Fault:	Not switching on - tried AA batteries		

Is the item contaminated?	Yes* ☐	No ☒	Don't know ☐
* State type of contamination: blood, body fluids, respired gases, pathological samples, chemicals (including cytotoxic drugs), radioactive material or any other hazard			
Has the item been decontaminated?	Yes(a) ☒	No(b) ☐	Don't know ☐
(a) What method of decontamination has been used? Please provide details:			
Cleaning:	☒		
Disinfection:	☒	Wiped all external aspects with ViruSolve wipe. disposed of CO2 sampling line	
Sterilisation:	☐		
(b) Please explain why the item has NOT been decontaminated:			

CONTAMINATED ITEMS SHOULD NOT BE RETURNED WITHOUT PRIOR AGREEMENT OF THE RECIPIENT

This item has been prepared to ensure safe handling and transportation:			
Name:	Lisa Angell	Position:	Head Anaesthesia RVN
Signature:	<i>Angell</i>		
Date:	lange@rvc.ac.uk	Tel:	lange@rvc.ac.uk

Please find enclosed unit.

On my initial conversation with Steve the unit would not power with the rechargeable battery or 4x new AA batteries.

Today (2 weeks later) the unit fired up with the unit's rechargeable battery but immediately powered down. Still no power with 4x AA's.

Many thanks

Lisa Angell