


bluepoint MEDICAL	Labeling Specifications	Rev. 3
Internal Document CONFIDENTIAL	Viamed Gas Sampling Line H (Sedana)	Page 1 of 7

Document number:	O-04-01-0003
-------------------------	---------------------

Labeling Specifications

Viamed Gas Sampling Line H (Sedana)

Authored	Function/ Position	Date	Signature
H. Fröhlich	Gas Exchange Monitors PLV	2024-10-14	
Reviewed	Function /Position	Date	Signature
C. George	Development Engineer	2024-10-14	
Reviewed and Approved	Function /Position	Date	Signature
B. Lindner	Gas Sampling Lines PLV	2024-10-14	

	Labeling Specifications	Rev. 3
Internal Document CONFIDENTIAL	Viamed Gas Sampling Line H (Sedana)	Page 2 of 7

Table of Content

1 Objective..... 3

2 Scope 3

3 Reference 3

4 Definitions and Abbreviations..... 3

5 Box label..... 4

 5.1 Single package label.....5

6 Hang tag label..... 6

7 Document History 7

	Labeling Specifications	Rev. 3
Internal Document CONFIDENTIAL	Viamed Gas Sampling Line H (Sedana)	Page 3 of 7

1 Objective

This document summarizes the product and packaging labels of the Viamed Gas Sampling Line H. This summary refers to the detailed label specifications of all labels used for manufacturing the Viamed Gas Sampling Line H.

2 Scope

This label summary applies to products listed below:

REF	Product name
8090121313V	Viamed Gas Sampling Line H, disposable, 2.5m/8ft

As updated according to design change CR-06-00-0001.

3 Reference

ID	Document No.	Rev.	Name/ Beschreibung
[R1]	n/a	3	8090121313V_Viamed_2,5m_Sampling_Line_H_box
[R2]	n/a	4	8090121313V_Viamed_2,5m_Sampling_Line_H_single
[R3]	I240910-022	1	I240910-022_17x57,15_Typenschild_72h_1korrr

4 Definitions and Abbreviations

Term	Description
REF	Product reference (order number)

	Labeling Specifications	Rev. 3
Internal Document CONFIDENTIAL	Viamed Gas Sampling Line H (Sedana)	Page 6 of 7

6 Hang tag label

Specifications:

Parameter	Specification
Label material	PE
Background color	Yellow (HKS 4 K)
Symbols and text color	Black
Label size	17 x 57mm
Hole size	6mm diameter
Drawing	[R3]



Figure 5: Hang tag label content



Figure 6: Hang tag label position

	Labeling Specifications	Rev. 3
Internal Document CONFIDENTIAL	Viamed Gas Sampling Line H (Sedana)	Page 7 of 7

7 Document History

Rev.	Approval date	Author	Change
0	2021-03-04	R. Jürgensen	Initial version
1	2021-10-14	R. Jürgensen	Added Rx symbol
2	2022-04-12	R. Jürgensen	Update labels to versions: Box label: B075-10/2021 Rev.2 Single label: A075-11/2021 Rev.3
3	2023-10-14	H. Fröhlich	- Remove body number at the CE (class I product). - Update label revision - Move MD symbol away from UDI barcode <u>Only box label:</u> - Remove redundant humidity and pressure symbol - Unify languages with IFU (add “ar”, “bg” and “sl”. Remove “lv”, “lt” and “sk”) <u>Only single label:</u> - add “Distributed by” to Viameds address as on box label. <u>Only hang tag label:</u> - Update material to PE (before PVC) - Increase label size to 17 x 57mm (before 17 x 55mm) to meet suppliers minimal size specification