

25850

# Service Repair Sheet SRS69016

Contact Name Jamie Smith  
 Company/ Hospital Name Pinderfields General Hospital  
 Department Medical Physics  
 Position Apprentice Clinical Engineer  
 Direct Phone 01924541050  
 General Phone 01924 201688  
 Opera Account 00005190  
 Email jamie.smith56@nhs.net  
 Order Number  
 Date Received 03/Feb/2025  
 Booked in By Robert Connor  
 Main Company Viamed  
 Type Return Quote

## VIAMEDclean

### Goods In Only

Decontamination  
 certificate provided  
 by customer ☒

Cleaned by Viamed,  
 if no declaration certificate  
 from customer ☒

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

### Goods Out Only

Cleaned by Viamed before  
 returning to customer

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

Notes 03/Feb/2025 Robert Connor

03/Feb/2025 Robert Connor

Returned 1 x MIPM TOF3D s/n T3D202101026, screen is cracked, includes full patient lead, cannot determine if batteries are resent, compartment sealed.

Ready For quote

CGneen 18:3:25

Repair Complete Signed \_\_\_\_\_

| SRN      | Equipment | Stock Ref | Serial Number | Warranty |
|----------|-----------|-----------|---------------|----------|
| SRN37599 | Microstim | 2510091   | T3D202101026  | N        |

2540100 x 1 @ £95 @ 1/2hr £47.50  
 S/N, SRS, SEN

2540101 x 1 £625 → given to me by Steve Nixon  
 S/N, SRS, SEN

UPS x 1 £12.