

Service Repair Sheet 69055

Contact Name	Christopher Courage
Company/ Hospital Name	Royal Cornwall Hospital
Department	Department of Clinical Technology
Position	Entry-Level Technician
Direct Phone	01872252488
Email	christopher.courage@nhs.net
Order Number	24027849
Date Received	10/Mar/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37680	1410000	GE71743920	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
