

Service Repair Sheet 69045

Contact Name	David Watson
Company/ Hospital Name	Darlington Memorial Hospital
Department	Clinical Engineering Department
Position	Clinical Engineering Technician
Direct Phone	01325743048
General Phone	01325380100
Opera Account	00001350
Email	david.watson28@nhs.net
Order Number	CDD0006276086
Date Received	10/Mar/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37681	0310030	VMD107	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		

Tom Thumb TT480 VMD107 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN37682	0310034	VMDH10	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		
	Qty: 1	0330218		

Tom Thumb TT490 VMDH10 has been stripped and cleaned then reassembled with new O rings and a new oxygen hose. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN37683	0310097	0401255	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		

Tom Thumb TT480 0401255 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.
