

Service Repair Sheet 69036
Contact Name Andrew Ruddick
Company/ Hospital Name Sunderland Royal Hospital
Position Clinical Engineer
Direct Phone 01915699933
General Phone 01915699933
Opera Account 00004992
Email andrew.ruddick@nhs.net
Order Number RLNP400117771
Date Received 10/Mar/2025
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37679	1410000	GE71743924	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.
