

Certificate of Calibration

Certificate number: 1743590859

V1000 Foetal Heart Simulator

Part Number: 1410000

Serial number: PR02913A10

This certificate reports recorded values for the instrument specified above.

The output rates were measured in accordance with procedure:
VM3COP55.01.

Equipment used: Sonicade FM6 foetal monitor (CE051)

TTi TF930 3GHz counter (CE185) - calibration
traceable to UKAS calibration standards.

V1000 BPM setting	Period values stated in milliseconds				Result PASS/FAIL
	Target	Lower limit	Upper limit	Actual	
60	1000.00	990.00	1010.00	993.11	PASS
90	666.70	660.03	673.37	665.63	PASS
120	500.00	495.00	505.00	500.03	PASS
150	400.00	396.00	404.00	400.13	PASS
180	333.30	329.97	336.63	332.64	PASS
210	285.70	282.84	288.25	285.83	PASS

Name: Phil Crossley

Date: 2nd April 2025

Valid Until 1st April 2026

Signed: _____