

Service Repair Sheet SRS69071

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

James Page

St James University Hospital

Medical Physics & Engineering

Medical Electronics Technician

0113 206 6882

01132433144

00002570

james.page3@nhs.net

24/Mar/2025

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: RC

Date: 24/3/25

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 18/Mar/2025 Steve Hardaker

18/Mar/2025 Steve Hardaker

Returning a Tom Thumb for service, VMDF25. Requires 3m hose replacement.

24/Mar/2025 Robert Connor

Received 1 x Tom thumb s/n VMDF25

Ready For quote

CGreen 26.3.25

Repair Complete Signed

CGreen 3.4.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37725	Tom Thumb	0310034	VMDF25	N

~~0308~~
03800000 x 1 @ £95
S/N, SRS, SRN

0330218 x 1 @ £45.70
SRS, SRN

UPS x 1 @ £12.