

Service Repair Sheet 69073

Contact Name	Karl Green
Company/ Hospital Name	Alexandra Hospital
Department	Technical Services
Direct Phone	01562823424
General Phone	01527 503030
Opera Account	00004295
Email	karl.green1@nhs.net
Date Received	31/Mar/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type
SRN37734	1410000	PR02913A10	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
