

CHECKED BY *C. Wans* DATE *02/04/25*
 RECEIVER # *311185* OCR # *multiple*
 WO/PO # *PD12500287/1*
 JFD PART NUMBER *DB200381* QTY *39*
 TRACEABLE MARK *AS HIGHLY LIGHTED*
107126 thru 107164



DECLARATION OF CONFORMITY

Description: **DB200381 Oxygen Sensor**
 Part number: **8010050**

Manufacturer: **Vandagraph Sensor Technologies Ltd.**
15 Station Road
Cross Hills
Keighley
West Yorkshire BD20 7DT
United Kingdom

Specification: **OEM Specification Datasheet – 8010050, v2.0 dated 6th May 2019**

Standards: **ISO 9001-2015**

JFD purchase order number: 12500287

Serial number/s	Manufacture – Year/Month
107126– 107164 (qty 39)	2025/03

I, the undersigned, hereby declare that the oxygen sensors listed above conform to the stated specifications.

Signature:

Name:

S Nixon
S Nixon

Position:

Director

Date:

20th March 2025



Invoice number: RST154533-1
Shipment date: 20th March 2025
JFD reference: 12500287

**Magnetic signature analysis of production oxygen sensors
through JFD QA goods-in procedures**

39 x 8010050 Oxygen Sensor – JFD DB200381

Once the analysis has been carried out, please complete the following and email back to VST:

info@vandagraphst.com

Result (PASS/FAIL):

QTY 39 PASS LST

Name: Calum Walls
Position: Quality Assurance Engineer.

Signed by:
On behalf of JFD



Date: 02 APR 2025

Vandagraph Sensor Technologies Limited
15 Station Road, Cross Hills, Keighley, West Yorkshire BD20 7DT, United Kingdom
Tel +44 (0) 1535 634900 Fax +44 (0) 1535 635582 e-mail: info@vandagraphst.com
www.vandagraphst.com

Company registered in England, No. 03116432

Purchase order

12500708

Date	Part number	Part number	Serial number	Output	Zero offset voltage	
	VST	JFD		mV	μ V	
20/03/2025	8010050	DB200381	107126	7.8	49	3.41
20/03/2025	8010050	DB200381	107127	7.0	50	3.19
20/03/2025	8010050	DB200381	107128	8.1	68	4.11
20/03/2025	8010050	DB200381	107129	7.8	57	3.56
20/03/2025	8010050	DB200381	107130	7.9	52	4.29
20/03/2025	8010050	DB200381	107131	7.9	62	3.57
20/03/2025	8010050	DB200381	107132	7.4	52	2.86
20/03/2025	8010050	DB200381	107133	7.8	81	3.85
20/03/2025	8010050	DB200381	107134	7.1	58	4.03
20/03/2025	8010050	DB200381	107135	7.6	63	1.98
20/03/2025	8010050	DB200381	107136	7.9	67	2.67
20/03/2025	8010050	DB200381	107137	7.0	92	4.07
20/03/2025	8010050	DB200381	107138	7.0	63	4.02
20/03/2025	8010050	DB200381	107139	7.4	76	2.43
20/03/2025	8010050	DB200381	107140	7.6	54	2.95
20/03/2025	8010050	DB200381	107141	7.5	43	3.27
20/03/2025	8010050	DB200381	107142	8.1	87	4.01
20/03/2025	8010050	DB200381	107143	8.0	29	2.56
20/03/2025	8010050	DB200381	107144	7.8	91	2.13
20/03/2025	8010050	DB200381	107145	7.2	58	3.17
20/03/2025	8010050	DB200381	107146	7.8	53	2.81
20/03/2025	8010050	DB200381	107147	8.1	58	2.53
20/03/2025	8010050	DB200381	107148	8.0	46	4.26
20/03/2025	8010050	DB200381	107149	8.1	62	3.75
20/03/2025	8010050	DB200381	107150	8.1	49	4.21
20/03/2025	8010050	DB200381	107151	7.8	53	4.01
20/03/2025	8010050	DB200381	107152	7.0	62	2.65
20/03/2025	8010050	DB200381	107153	7.1	82	3.33
20/03/2025	8010050	DB200381	107154	7.9	58	2.54
20/03/2025	8010050	DB200381	107155	7.4	55	4.08
20/03/2025	8010050	DB200381	107156	8.1	66	3.33
20/03/2025	8010050	DB200381	107157	7.9	55	2.56
20/03/2025	8010050	DB200381	107158	8.1	54	4.81
20/03/2025	8010050	DB200381	107159	8.1	75	4.22
20/03/2025	8010050	DB200381	107160	8.0	82	3.21
20/03/2025	8010050	DB200381	107161	7.6	59	3.21
20/03/2025	8010050	DB200381	107162	8.1	64	2.41
20/03/2025	8010050	DB200381	107163	7.7	57	4.43
20/03/2025	8010050	DB200381	107164	8.0	59	2.68

 < 5nT
 > 5nT