

Service Repair Sheet 69083

Contact Name	Trudy Moule
Company/ Hospital Name	County Hospital
Department	Clinical Technology
Position	Clinical Technology Admin & Helpdesk Support
Direct Phone	01785230305
General Phone	01785257731
Opera Account	00004840
Email	Trudy.Moule@uhnm.nhs.uk
Order Number	X029379
Date Received	02/Apr/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37738	1410000	PR0495A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
