

Service Repair Sheet 69090

Contact Name	Gillian Nixon
Company/ Hospital Name	South Tyneside District Hospital
Department	Clinical Engineering Department
Position	Clinical Engineering Administrator
Direct Phone	0191 404 1160
General Phone	0191 404 1000
Opera Account	00004710
Email	gillian.nixon3@nhs.net
Order Number	RLNP400119954
Date Received	03/Apr/2025
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN37741	0310200	BX2260065	Blender	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380020		
	Qty: 1	0332000		

Blender BX2260065 has had an annual service with new gas filters and the calibration has been checked

---