

Service Repair Sheet SRS69016

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Jamie Smith
Pinderfields General Hospital
Medical Physics
Apprentice Clinical Engineer
01924541050
01924 201688
00005190
jamie.smith56@nhs.net

03/Feb/2025
Robert Connor
Viamed
Quote

VIAMEDclean

Goods In Only
Decontamination
certificate provided
by customer



Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____

Date: _____



Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 03/Feb/2025 Robert Connor

03/Feb/2025 Robert Connor

Returned 1 x MIPM TOF3D s/n T3D202101026, screen is cracked, includes full patient lead, cannot determine if batteries are resent, compartment sealed.

Ready For quote

CGreen 18.3.25

Repair Complete Signed

CGreen 10.4.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37599	Microstim	2510091	T3D202101026	N

2540100 x 1 @ £95 @ 1/2hr £47.50
S/N, SRS, SEN

2540101 x 1 £625
~~S/N~~, SRS, SEN

→ given to me by Steve Nixon

UPS x 1 £12.