

Service Repair Sheet 69016

Contact Name	Jamie Smith
Company/ Hospital Name	Pinderfields General Hospital
Department	Medical Physics
Position	Apprentice Clinical Engineer
Direct Phone	01924541050
General Phone	01924 201688
Opera Account	00005190
Email	jamie.smith56@nhs.net
Date Received	03/Feb/2025
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type
SRN37599	2510091	T3D202101026	Microstim

Complete - Repaired  
Time :0.5 Hour(s)

Parts Replaced

Qty: 1	2540101
Qty: 0.5	2540100

The TOF3D has been repaired. The main PCB and display have been replaced. The unit has been tested and is working as expected.

---