

Germany  
90181990-00

# Service Repair Sheet SRS69007

Contact Name  
Company/ Hospital Name  
Department  
Position  
Direct Phone  
General Phone  
Opera Account  
Email  
Order Number  
Date Received  
Booked in By  
Main Company  
Type Return

Mr Mahesha  
BDI India Private Limited  
Customer Support  
Customer Support  
918022560418  
918022560418  
00006454  
mahesha@bdi-india.com

05/Mar/2025  
Robert Connor  
Viamed  
Quote

<b>VIAMEDclean</b>	
<b>Goods In Only</b>	
Decontamination certificate provided by customer <input checked="" type="checkbox"/>	
Cleaned by Viamed, if no declaration certificate from customer <input type="checkbox"/>	
Signed: _____	_____
Date: _____	_____
<b>Goods Out Only</b>	
Cleaned by Viamed before returning to customer	
Signed: _____	_____
Date: _____	_____

Notes 20/Jan/2025 Kate Griffiths

20/Jan/2025 Kate Griffiths

10 V1000s to send in for calibration. I will ask as usual for the devices to be cleaned before they send them. Will also ask for cert of decont. status.

PR02380A12, PR02529A14, PR02529A18, PR02540A13, PR02540A14, PR02529A22, PR02529A15, PR02529A17, PR02529A11, PR02529A20

05/Mar/2025 Robert Connor

Received 10 x V1000 s/n PR02529A11, with grey cover and missing battery door, PR02529A14, with grey cover and missing battery door, PR02529A15, PR02529A17, with grey cover and missing battery door, PR02529A18, with 4 x AA battery, PR02529A20, with grey cover and missing battery door, PR02529A22, PR02380A12 missing battery door, PR2540A13 with grey cover and missing battery door, PR2540A14 with grey cover and missing battery door.

Units have not been decontaminated at goods in as buildup of dirt is too great for the cleaning products available. Units need deep clean before testing.

Ready For quote

CGreen 11.3.25

USD DIST.

Repair Complete Signed

CGreen 10.4.25

Refused by customer  
April 25.  
CG

SRN	Equipment	Stock Ref	Serial Number	Warranty
* SRN37669	Foetal Simulator	1410000	PR02529A11	
SRN37670	Foetal Simulator	1410000	PR02529A14	
SRN37671	Foetal Simulator	1410000	PR02529A15	
* SRN37672	Foetal Simulator	1410000	PR02529A17	
SRN37673	Foetal Simulator	1410000	PR02529A18	
SRN37674	Foetal Simulator	1410000	PR02529A20	
SRN37675	Foetal Simulator	1410000	PR02529A22	
SRN37676	Foetal Simulator	1410000	PR02380A12	
* SRN37677	Foetal Simulator	1410000	PR02540A13	
SRN37678	Foetal Simulator	1410000	PR02540A14	

\* Added 21.3.25  
1430276  
SRS 69007  
SRN, 37669-37670  
37672, 37674  
37676-37678  
x7 @ \$58.95

14800000 x 10 @ \$63.40  
S/N's SRS, SRN's

1430309 x 10 FCC as part of service  
SRS, SRN's

1430288 x 3 @ \$5.50 \*  
SRS, SRN's

14400001 x 1 @ \$66.  
for cleaning.  
BC @ \$35  
WPS x 1  
36x36x36cm  
Seg.  
\$128.73  
NOW EXW CG  
10.4.25