

Service Repair Sheet SRS69070

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Sudath Weeraman

Prime Lanka Associate

9410718732285

00006714

sudathweeraman17@gmail.com

14/Apr/2025

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 17/Mar/2025 Kate Griffiths

17/Mar/2025 Kate Griffiths

Customer has a R-42v with ``calibration failure``. s/n: V101777 under warranty

14/Apr/2025 Robert Connor

Received 1 x R42 s/n V101777.

Ready For quote

Repair Complete Signed

Cgreen 15.4.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37758	Oxygen Sensor	0110042	V101777	Yes

0140013 x 1 FOC
S/N, SRS, SRN

0110042 x 1 FOC
New S/N, SRS, SRN

EXW Skycargo.
23x12x12cm 0.10kg