

Service Repair Sheet 69105

Contact Name	Robert Barr
Company/ Hospital Name	Cardiotec
Direct Phone	07980656316
General Phone	07980656316/02895908890
Opera Account	CID17049
Email	robert@cardiotec.co.uk
Order Number	PO-1820
Date Received	17/Apr/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37763	1410000	PR03372A17	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
