

Service Repair Sheet 69110  
Contact Name Purchasing  
Company/ Hospital Name Eden Medical (UK) Ltd  
Direct Phone 01314406500  
General Phone 01314406500  
Opera Account 00012178  
Email info@edenmedical.co.uk  
Date Received 28/Apr/2025  
Booked in By Robert Connor

---

| Repair         | Ref     | S/N        | Equipment Type   |                                        |
|----------------|---------|------------|------------------|----------------------------------------|
| SRN37812       | 1410000 | PR02643A11 | Foetal Simulator | Complete - Repaired<br>Time :0 Hour(s) |
| Parts Replaced |         |            |                  |                                        |
|                | Qty: 1  | 1480000    |                  |                                        |
|                | Qty: 1  | 1430309    |                  |                                        |

V1000 has had transducer interface cushion replaced and calibration has been checked.

---