



Kate Griffiths <viamed.kate.griffiths@gmail.com>

RE: FW: Oxygen Sensor - záruka?

2 messages

Oskar Šlechta <o.slechta@chromservis.eu>

30 April 2025 at 11:56

To: "kate.griffiths@viamed.co.uk" <kate.griffiths@viamed.co.uk>

Cc: "office@viamed.co.uk" <office@viamed.co.uk>

Dear Kate,

unfortunately, we received a new claim for **R-22A sensor, s/n 169712 F4**, received in December 2024 under our order DZ2024-529 (we ordered 60 sensors together under this order)..

After installation, the value was 0,9mV only. This sensor doesn't work. I would like to ask you for start of claim process. Of course, we'll send faulty sensor you to back.

Thanks so much, best regards

Oskar

Oskar Šlechta

Senior Internal Sales Advisor

Chromservis Praha

[Jakobiho 327](#)[CZ-10900 Praha 10](#)

Czech Republic

Tel.: +420 274 021 221

Fax: +420 274 021 220

E-mail: o.slechta@chromservis.euwww.chromservis.eu

Kate Griffiths <kate.griffiths@viamed.co.uk>

30 April 2025 at 14:35

Draft To: Oskar Šlechta <o.slechta@chromservis.eu>

Bcc: Main Account <viamedinbox@gmail.com>