

Service Repair Sheet 69087

Contact Name	Dane Milne
Company/ Hospital Name	University Hospitals Plymouth NHST
Department	Clinical Engineering
Position	Clinical Engineering IT Systems Administrator
Direct Phone	01752433595
General Phone	01752202082
Opera Account	00004127
Email	dane.milne@nhs.net
Date Received	16/Apr/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN37762	1410000	PR03605A13	Foetal Simulator	Complete - Repaired
Parts Replaced				Time :0 Hour(s)
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
