

Service Repair Sheet SRS69115

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Oskar Slechta

Chromservis s.r.o

Senior Internal Sales Advisor

420274021221

420274021221

00012607

o.slechta@chromservis.eu

12/May/2025

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☐

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 30/Apr/2025 Kate Griffiths

30/Apr/2025 Kate Griffiths

R-22A sensor to be returned under warranty. S/n: 169712. The value was 0.9mV only

12/May/2025 Robert Connor

Received 1 x R22A s/n 169712

Ready For quote

Repair Complete Signed

CCGreen 13.5.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37846	Oxygen Sensor	0110122	169712	Yes

0140013 x 1 FOC
old
S/N, SRS, SRN.

0110122 x 1 FOC
New S/N, SRS, SRN

UPS x 1 FOC.