

Service Repair Sheet SRS69123

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

David Edmondson

Royal Preston Hospital

Medical Engineering

Technician

01772 524598

01772716565

00004260

david.edmondson@lthtr.nhs.uk

12/May/2025

Robert Connor

Viamed

Quote

**VIAMEDclean**

**Goods In Only**

Decontamination  
certificate provided  
by customer ☒

Cleaned by Viamed,  
if no declaration certificate  
from customer ☐

Signed: \_\_\_\_\_  
Date: \_\_\_\_\_

**Goods Out Only**

Cleaned by Viamed before  
returning to customer

Signed: \_\_\_\_\_  
Date: \_\_\_\_\_

Notes 12/May/2025 Robert Connor

12/May/2025 Robert Connor

Returned 1 x tom thumb s/n VMDI25, 1 x blender s/n BX1551978. Tom thumb gauge is not tightly screwed on, can be undone by hand.

Ready For quote

Repair Complete Signed

CCGreen 13-5-25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37844	Tom Thumb	0310030	VMDI25	N
SRN37845	Blender	0310200	B1551978	N

0380000 x 1  
S/N, SRS, SRN

0380020 x 1  
S/N, SRS, SRN

update CVM154936  
on a call off  
service order  
cg 13-5-25