Service Repair Sheet SRS69126

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Hussein Remtulla

St James University Hospital

Clinical Engineering

Medical Electronics Technician

01132066882

01132433144

00002570

hussein.remtulla@nhs.net

13/May/2025

Robert Connor

Viamed

Quote

| VIAMEDCIES Goods In Only Decontamination certificate provided by customer | an 🗵 |
|---|------|
| Cleaned by Viamed, If no declaration certificate from customer Signed: | |
| Goods Out Only Cleaned by Viamed beforeturning to customer Signed: Date: | ore |

Notes 12/May/2025 Kate Griffiths 12/May/2025 Kate Griffiths

Sending in a DB3. Fault: the LED does not light up every time button is pressed - intermittent. The signal/current goes through the leads but the LED at times stays off. S/N: H0002351. Asked for cert of decontamination and batteries removed 13/May/2025 Robert Connor

Received 1 x microstim DB3 s/n H0002351, with patient lead and 9v battery

Ready For quote

careen 145-25

Repair Complete Signed

| SRN | Equipment | Stock Ref | Serial Number | Warranty |
|----------|---------------|-----------|---------------|----------|
| SRN37853 | Microstim DB3 | 2510000 | H0002351 | 2 |

2540010 × 1 @ £65 SIN, SRS, SRN

UPS + 1 2 E12.