

Service Repair Sheet 69092

Contact Name	Pierce Laurens
Company/ Hospital Name	Queen Elizabeth The Queen Mother Hospital
Department	EME Department
Position	Senior EME Technician
Direct Phone	01843234346
General Phone	01843225544
Opera Account	00003665
Email	pierce.laurens@nhs.net
Order Number	44009944
Date Received	08/Apr/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37749	1410000	PR0542A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
