

Calibration Certificate

The VN202 Oxygen analyser has been tested for accuracy by using a UKAS certified voltage source.

Calibrated voltage source	Microcal
ISO calibration index number	CE076
Serial number of the Microcal	6939K6
Certificate number of the Microcal	1895708
Serial number of the VN202	00822
Certificate number of the VN202	1747479274

Voltage input:	Target reading:	VN202 reading:
5.0mV	10%	9.9%
10.45mV	20.9%	20.9%
10.5mV	21%	20.9%
12.5mV	25%	24.9%
25.0mV	50%	50.0%
37.5 mV	75%	75.0%
50.0mV	100%	100%

Tested by: M.Green

Date: 17/05/2025

Signed: