

Service Repair Sheet SRS69133

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Lesley McSporran

Dumfries & Galloway Royal Infirmary

Medical Physics Department

Medical Engineering Technician S

01387241816

01387246246

00001480

lesley.mcsporrn@nhs.scot

22/May/2025

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 22/May/2025 Robert Connor

22/May/2025 Robert Connor

Returned 2 x tom thumb s/n VMCC42, gauge cover loose, rail mounting clamp wingnut smashed, VMCC43 gauge face loose.

Ready For quote

CGNEEN 23.5.25

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37878	Tom Thumb	0310030	VMCC43	22
SRN37879	Tom Thumb	0310030	VMCC42	22

03800000 x 2 @ £95 each
S/N, SRS, SRN's

0330203 x 2 @ £128.30 each
~~SRN~~, SRS, SRN's

0330212 x 2 @ £3.10 each
SRS, SRN's

UPS x 1 @ £12