

Service Repair Sheet 69134

Contact Name	John Itter
Company/ Hospital Name	Salisbury District Hospital
Department	Medical Device Management Services
Position	Senior Technician
Direct Phone	01722336262
General Phone	01722 33 62 62
Opera Account	00004470
Email	john.itter@nhs.net
Order Number	351124508
Date Received	02/Jun/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37884	1410000	GE71743923	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
