

Service Repair Sheet SRS69135

Contact Name Luke Banks
 Company/ Hospital Name Luke Banks
 Department
 Position
 Direct Phone 07770235256
 General Phone 07770235256
 Opera Account CID29122
 Email admin@elroyairmed.com
 Order Number DVM150164-1
 Date Received 02/Jun/2025
 Booked in By Robert Connor
 Main Company Viamed
 Type Return Warranty

VIAMEDclean
Goods In Only
 Decontamination certificate provided by customer ☒

Cleaned by Viamed, if no declaration certificate from customer ☒
 Signed: RC
 Date: 2/6/25

Goods Out Only
 Cleaned by Viamed before returning to customer
 Signed: _____
 Date: _____

Notes 27/May/2025 Aqib Majeed

27/May/2025 Aqib Majeed

returning a Max02 Me Monitor SN:JJ60899011 as it has a screen fault, still under warranty

02/Jun/2025 Robert Connor

Received 1 x MAXO2ME s/n JJ60899011, with coiled cable 2375067, MAX550E s/n JJ54299185, flow diverter, 4 x AA battery.

Ready For quote

Repair Complete Signed

CCNOR 5.6.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37885	Oxygen Monitor	0111263	JJ60899011	Yes
SRN37886	Oxygen Monitor Cable	0131700	Tracking20	
SRN37887	Oxygen Sensor	0110452	JJ54299185	Yes

3

0140050
~~0140000~~ x 1 FOC
 old S/N, SRS, SRN
 See below for replacement

0111263 x 1 FOC
 New S/N, SRS, SRN

0140011 x 1 FOC
 S/N, SRS, SRN

0140025 x 1
 S/N, SRS, SRN

UPS x 1 FOC.