

Service Repair Sheet 69141

| | |
|------------------------|-----------------------------|
| Contact Name | Arlene Gallagher |
| Company/ Hospital Name | Forth Valley Royal Hospital |
| Department | Medical Physics Dept |
| Position | Ms Gallagher |
| Direct Phone | 01324 566953 |
| General Phone | 01324 566000 |
| Opera Account | 00004880 |
| Email | arlene.gallagher@nhs.scot |
| Order Number | VMF15441920 |
| Date Received | 02/Jun/2025 |
| Booked in By | Robert Connor |

| | | | | |
|----------------|---------|------------|------------------|--|
| Repair | Ref | S/N | Equipment Type | |
| SRN37883 | 1410000 | PR02724A15 | Foetal Simulator | Complete - Repaired Time :0 Hour(s) |
| Parts Replaced | | | | |
| | Qty: 1 | 1480000 | | |
| | Qty: 1 | 1430309 | | |

V1000 has had transducer interface cushion replaced and calibration has been checked.
