

Service Repair Sheet 69099

Contact Name	Kamal Pacitti
Company/ Hospital Name	North Middlesex University Hospital
Department	Clinical Engineering Department
Position	Specialist Clinical Engineer
Direct Phone	02088873790
General Phone	02088872000
Opera Account	00003070
Email	kamal.pacitti@nhs.net
Order Number	RFL626714
Date Received	24/Apr/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37781	0310200	BX1854196	Blender	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	0380019	
		Qty: 1	0332001	

Blender BX1854196 has been stripped and cleaned. All O rings, diaphragms etc have been replaced and it has been calibrated.

Repair	Ref	S/N	Equipment Type	
SRN37782	0310200	BX1854639	Blender	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	0380019	
		Qty: 1	0332001	

Blender BX1854639 has been stripped and cleaned. All O rings, diaphragms etc have been replaced and it has been calibrated.

Repair	Ref	S/N	Equipment Type	
SRN37783	0310200	BX1855845	Blender	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	0380019	
		Qty: 1	0332001	
		Qty: 1	0332009	

Blender BX1855845 has been stripped and cleaned. The O2 NIST inlet has been replaced due to flaking chrome. All O rings, diaphragms etc have been replaced and it has been calibrated.

