

Service Repair Sheet 69139

Contact Name	Andrew Brown
Company/ Hospital Name	North Manchester General Hospital
Department	Materials Management Supervisor
Direct Phone	01619223735
General Phone	01617954567
Opera Account	00003580
Email	Andrew.Brown3@mft.nhs.uk
Order Number	000515345
Date Received	03/Jun/2025
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Serviced
SRN37934	2510000	H0001373	Microstim DB3	Time :0 Hour(s)
Parts Replaced				
	Qty: 1	2540010		
	Qty: 1	2520000		

The Microstim DB3 has been serviced and the patient lead has been replaced.

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN37935	2510000	M0005512	Microstim DB3	Time :0 Hour(s)
Parts Replaced				
	Qty: 1	2540010		
	Qty: 1	2520000		

The Microstim DB3 has been repaired. The PCB and patient lead have been replaced.
