

Service Repair Sheet 69154

Contact Name	Simon Pugh
Company/ Hospital Name	Princess Royal Hospital
Department	Medical Engineering Services
Position	Medical Engineer
Direct Phone	01952 641222 EXT 4262
General Phone	01743261149
Opera Account	00005117
Email	simon.pugh3@nhs.net
Order Number	51248448
Date Received	12/Jun/2025
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN37949	1410000	PR0773A14	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		
	Qty: 0.25	1440000		
	Qty: 1	1430297		

V1000 has had transducer interface cushion and potentiometer replaced. The calibration has been checked.

---