

Service Repair Sheet 69157
Contact Name David Nunn
Company/ Hospital Name Sunderland Royal Hospital
Direct Phone 01915656256
General Phone 01915656256
Opera Account 00004992
Email david.nunn2@nhs.net
Order Number RLNP400123889
Date Received 23/Jun/2025
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37958	1410000	PR02010A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
		Qty: 1	1480000	
		Qty: 1	1430309	

V1000 has had transducer interface cushion replaced and calibration has been checked.
