Service Repair Sheet 69166

Contact Name Nathaniel Ford

Company/ Hospital Name Isle of Wight NHS Primary Care Department Medical Electronics Department

Position Medical Devices Coordinator & Senior Technician

 Direct Phone
 01983822099

 General Phone
 01983822099

 Opera Account
 00004205

Email nathanielford@nhs.net

Order Number 282178539
Date Received 02/Jul/2025
Booked in By Robert Connor

Repair Ref S/N Equipment Type

SRN37968 1410000 PR03072A10 Foetal Simulator Complete - Repaired

Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000 Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.